MERCATOR MEDICAL



nitrylex[®] beFree long

The instruction below should be used in conjunction with detailed information on the packaging.

Short description of the product

Nitrile examination and protective gloves, powder-free, non-sterile for disposable use

Full description of the product		
Raw material	: nitrile	
External surface	: microtextured with fingertip textured,	
	polymerized	
Internal surface	: polymerized + chlorinated	
Cuff	: beaded	
Colour	: blueberry	
Shape	: ambidextrous, fitting to the right and left hand	
Size range	: XS (5-6), S (6-7), M (7-8), L (8-9), XL (9-10)	
AQL	: 1.0	
Quantity in packaging	: 100 pcs. by weight	
Shelf life	: 3 years (from the date of manufacturing)	

Storage instructions

It is recommended to store the gloves in dry place, in the temperature of $5-35^{\circ}$ C and to protect them against direct sunlight and fluorescent light. Recommended relative humidity in the room where the gloves are stored is 60 ±20%.

Keep the gloves in a distance of not less than 1m from heating devices, sources of fire and ozone.

Do not keep in direct vicinity of solvents, oils, fuels and lubricants.

Food contact

Gloves are marked with food contact symbol \times and comply with the requirements of Regulation (EU) No 10/2011, European Regulation (EC) No 1935/2004 and with Regulation (EC) No 2023/2006 on Good Manufacturing Practice. Gloves are suitable for handling any type of food and have been tested for Overall Migration Test acc. EN 1186:

Extraction conditions (tested for 1 h in 40°C)	Analysis results [mg/dm ²]	Test Result (limit < 10 mg/dm²)
50% ethanol	4.6	Pass
3% acetic acid	7.6	Pass
Vegetable oil	<2.0	Pass
Tenax	<2.0	Pass

MDR classification & compliance

Gloves are classified as class I according to Annex VIII of the Regulation (EU) 2017/745 and comply to standards:

EN 455-1:2000, EN 455-2:2015, EN 455-3:2015, EN 455-4:2009, EN ISO 15223-1:2016, EN 1041:2008+A1:2013.


PPER classification & compliance

Gloves are category III Personal Protective Equipment as per Annex I of the Regulation 2016/425 and comply to standards:

EN 420:2003+A1:2009, EN ISO 374-1:2016 (Type C), EN 374-2:2014, EN 16523-1:2015, EN 374-4:2013, EN ISO 374-5:2016.

Declaration of Conformity can be found under below web address: <u>https://mercatormedical.eu/products/gloves/examination-and-protective-gloves/nitrylex-befree-long</u>

EU Type Examination (Module B) Notified Body: Notified Body 2777 Satra Technology Europe Ltd Bracetown Business Park, Clonee, Dublin 15



CE0598 Notified Body 0598 SGS FIMKO OY P.O. Box 30 (Särkiniementie 3) 00211 Helsinki, Finland

Intended use

These are non-sterile examination and protective gloves for single use, intended for use in medical field to: protect patient and user from crosscontamination, conducting medical examinations, diagnostic and therapeutic procedures and for handling medical contaminated material. Gloves are classified as Medical Devices Class I and as a Personal Protective Equipment Category III. Their design and labelling corresponds to the requirements of the European Regulation 2017/745 on Medical Device and the European Regulation 2016/425 on Personal Protective Equipment. Gloves should be used solely according to their intended application.

Precautions and indications for use

Dry hands before putting the gloves on. Before usage, inspect the gloves for any defect or imperfections. Use at least 1 pair of gloves for one patient and one procedure, these are disposable gloves. Do not let chemical substances get under the gloves through the cuff. If a chemical substance reaches the skin, wash it away immediately with plenty of water with soap. If the gloves get punctured, torn or broken during their use, take them off and put on the new ones. Avoid using gloves dirty in the inside as they may cause irritation leading to skin inflammation or more serious damages. The gloves should not be used in contact with open fire and to protect against any sharp tools. The gloves are not intended for welding, electric shock protection, ionizing radiation or from the effect of hot or cold objects.

This information does not reflect the actual duration of protection in the workplace and the differentiation between mixtures and pure chemicals.

The chemical resistance has been assessed under laboratory conditions from samples taken from the palm only (except in case where glove is equal to or over 400 mm – where the cuff is tested also) and relates only to the chemical tested and to the tested specimen. It can be different if the chemical is used in a mixture.

The penetration resistance has been assessed under laboratory conditions and relates only to the tested specimen.

It is recommended to check that the gloves are suitable for the intended use because the conditions at the workplace may differ from the type test depending on the temperature, abrasion and degradation.

When used, protective gloves may provide less resistance to the dangerous chemical due to changes in physical properties. Movements, snagging, rubbing, degradation caused by the chemical contact etc. may reduce the actual use time significantly. For corrosive chemicals, degradation can be the most important factor to consider in selection of chemical resistant gloves.

Gloves are suitable for special purposes as they are examination gloves where risk of injury to the wrist is considered to be minimal, gloves are shorter than EN 420 min. length requirement.

Components / hazardous components

Some gloves may contain components known to be a possible cause of allergy for person allergic to them, who may develop contact irritation and/or allergic reaction. In case of an allergic reaction, seek medical assistance immediately.

Disposal


Used gloves can be contaminated with contagious or other hazardous substances. They should be disposed of in accordance with local regulation. Gloves should be buried or burned under controlled conditions.

Manufacturer


MERCATOR MEDICAL S.A. ul. H. Modrzejewskiej 30 31-327 Cracow, Poland www.mercatormedical.eu

Dublin, Ireland

MERCATOR MEDICAL



MERCATOR MEDICAL



■ HOW TO TAKE THE GLOVES OFF?

