



# EC CERTIFICATE

## Full Quality Assurance System Medical Devices Directive 93/42/EEC Annex II (Excluding Section 4)

Company Name : Aqua Medikal Tıbbi Araç ve Gereçler İnş. San. Dış Tic. Ltd. Şti.

Company Address : Esentepe Mah. 2953. Sk. 11 Blok No:54 Sultangazi İSTANBUL / TURKEY

Related Directives and Annex : 93/42/EEC Medical Devices Directive - Annex II (Excluding Section 4)

Product : - Non-Sterile 2% Gluteraldehyde High-Level Medical Device Disinfectant - Class IIb  
- Non-Sterile Alcohol-Containing Medical Device Disinfectant - Class IIa  
- Non-Sterile Lubricating Gel - Class IIa  
- Non-Sterile Lice Spray - Class IIa

GMDN : 40579, 58082, 33587, 45370

Certificate Number : M.2021.106.14611  
Report Number : MD.4149.IB  
Initial Assessment Date : 24.03.2021  
Registration Date : 21.05.2021  
Revision Date /No : -  
Expiry Date : 27.05.2024



UDEM International Certification  
Auditing Training Centre Industry  
and Trade Inc. Co.

UDEM hereby declares that the requirements of Annex II, excluding section 4 of the 93/42/EEC Directive have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance audits, defined by Annex II, section 5 of the forementioned directive. According to Annex II, section 4 an EC design- examination certificate is required for placing the Class III devices on the market. UDEM's responsibility for class I devices covered by the EC certificate is limited to manufacturing issues related to safeguarding and maintaining sterile conditions, if the device is sterile; and manufacturing issues related to product's conformity with metrological requirements, if it has measurement function. This certificate remains as the property of UDEM International Certification Auditing Training Centre Industry and Trade Inc. Co. to whom it must be returned upon request. The above named company and UDEM must keep a copy of this certificate for 5 years from the registration of the certificate. Usage of the CE mark is under the responsibility of the manufacturer with the completion of EC Declaration of Conformity. The above mentioned company must notify all changes related with the approved product to UDEM. If UDEM will not renew the expiry date of this certificate in question, the mentioned



**Address:** Mutlukent Mahallesi 2073 Sokak (Eski 93 Sokak) No:10 Çankaya – Ankara – TURKEY  
**Phone:** +90 0312 443 03 90 **Fax:** +90 0312 443 03 76  
**E-mail:** info@udemltd.com.tr www.udem.com.tr