

## Manufacturer's Declaration

in relation to Regulation (EU) 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices, in particular with respect to

- the validity of certificates issued under Council Directive 90/385/EEC on Active Implantable Medical Devices (AIMDD) or Council Directive 93/42/EEC on Medical Devices (MDD) (Directive Certificates) *and/or*<sup>1</sup>
- the compliance of the devices and us as their manufacturer with the conditions for the continued placing on the market and putting into service

|   |   |
|---|---|
| Manufacturer name                               | Guangzhou AXD Electronic Co., Ltd.  |
| Manufacturer address and contact details        | Factory 3, No. 9, Changxin Road,<br>Changshapu, Zhongluotan Town, 510550<br>Baiyun District, Guangzhou City, Guangong<br>Province, P.R. China |
| Single Registration Number (SRN) (if available) | CN-MF-000037466   |

|   |   |
|---|---|
| Authorised Representative name (if applicable)        | Shanghai International Holding Corp. GmbH<br>(Europe) |
| Authorised Representative address and contact details | Eiffestrasse 80, 20537, Hamburg, Germany              |
| Single Registration Number (SRN) (if available)       |   |

|   |  |
|---|--|
| Notified body name (if applicable)  | SGS Fimko Ltd.<br><input type="checkbox"/> See attached schedule |
| Notified body number (if applicable)  | CE0598<br><input type="checkbox"/> See attached schedule         |
| Directive Certificate number(s)<br>to which this confirmation is made (if applicable) | FI21/07010<br><input type="checkbox"/> See attached schedule     |
| Original expiry date as indicated on the Directive                                    | 24 <sup>th</sup> , May, 2024                                     |

<sup>1</sup> The first condition is not applicable in case of devices for which the conformity assessment procedure pursuant to MDD did not require the involvement of a notified body, for which the declaration of conformity was drawn up prior to 26 May 2021 and for which the conformity assessment procedure pursuant to this Regulation requires the involvement of a notified body.





Guangzhou AXD Electronic Co., Ltd.

Factory 3, No. 9, Changxin Road, Changshapu, Zhongluotan Town, 510550 Baiyun District, Guangzhou City, Guangdong Province, P.R. China

- We do not intend to lodge an application for conformity assessment by 26 May 2024, therefore the transition period will end on 26 May 2024.

- Expired/expires *after* 20 March 2023:

*Choose one applicable statement:*

Formal application(s) to the notified body in accordance with Section 4.3, first subparagraph of Annex VII MDR for conformity assessment has/have been made or will be made/submitted by us to a notified body no later than 26 May 2024 for the device(s) listed in the attached schedule or its/their substitute(s) and signed written agreement(s) is/will be in place in accordance with Section 4.3, second subparagraph of Annex VII MDR before 26 September 2024.

- We do not intend to lodge an application for conformity assessment by 26 May 2024, therefore the transition period will end on 26 May 2024.

➤ **Upclassified devices**

In case of devices for which the conformity assessment procedure pursuant to MDD did not require the involvement of a notified body, for which the declaration of conformity was drawn up prior to 26 May 2021 and for which the conformity assessment procedure pursuant to this Regulation requires the involvement of a notified body:

*Choose one applicable statement:*

Formal application(s) to the notified body in accordance with Section 4.3, first subparagraph of Annex VII MDR for conformity assessment has/have been made or will be made/submitted by us to a notified body no later than 26 May 2024 for the device(s) listed in the attached schedule or its/their substitutes and signed written agreement(s) is/will be in place in accordance with Section 4.3, second subparagraph of Annex VII MDR before 26 September 2024.

- We do not intend to lodge an application for conformity assessment by 26 May 2024, therefore the transition period will end on 26 May 2024.

➤ **Quality Management System (QMS)**

*Choose one applicable statement:*

A QMS in accordance with Article 10(9) MDR will be put in place by no later than 26 May 2024.

A QMS in accordance with Article 10(9) MDR is in place.

A notified body has issued the attached certificate for the MDR-compliant QMS.

➤ **Device(s) as listed in the attached schedule**

- The device(s) continue to comply with the AIMDD or MDD.
- There are no significant changes in the design and intended purpose.
- The device(s) do not present an unacceptable risk to health or safety of patients, users or other persons, or to other aspects of the protection of public health.



Guangzhou AXD Electronic Co., Ltd.

Factory 3, No. 9, Changxin Road, Changshapu, Zhongluotan Town, 510550 Baiyun District, Guangzhou  
City, Guangong Province, P.R. China

**Signed for and on behalf of the manufacturer:**

Guangzhou AXD Electronic Co., Ltd.

2024.4.18

General Manager: Mark Gao

Contact: axd@axd.net.cn



**Guangzhou AXD Electronic Co., Ltd.**

Factory 3, No. 9, Changxin Road, Changshapu, Zhongluotan Town, 510550 Baiyun District, Guangzhou City, Guangong Province, P.R. China

**Schedule of Devices**

The above Manufacturer's Declaration is valid for the following devices:

| <b>Identification of the device(s)<sup>3</sup></b><br>(e.g., device name, family/group name, device model or catalogue number) | <b>Directive Certificate number(s) to which this confirmation is made</b><br>(if applicable) | <b>Original expiry date as indicated on the Directive Certificate (s) prior to the extension of the validity</b><br>(if applicable) | <b>Notified Body name and number that issued the Directive Certificate</b><br>(if applicable) | <b>Notified Body name and number where the MDR application was lodged/contract signed</b><br>(if applicable) | <b>End date of extended validity / transition period</b> | <b>Substitute Device(s)</b><br>(if applicable) |
|--|--|---|---|--|--|--|
| Medical Compressor Nebulizer   | FI21/07010   | 24 <sup>th</sup> , May, 2024  | SGS Fimko Ltd. 0598   | SGS Fimko Ltd. 0598  | 31st, December, 2028                                     | /  |
| Digital Arm Blood Pressure Monitor   | FI21/07010   | 24 <sup>th</sup> , May, 2024  | SGS Fimko Ltd. 0598   | SGS Fimko Ltd. 0598  | 31st, December, 2028                                     | /  |
| Digital Thermometer  | FI21/07010   | 24 <sup>th</sup> , May, 2024  | SGS Fimko Ltd. 0598   | SGS Fimko Ltd. 0598  | 31st, December, 2028                                     | /  |
| Infrared Thermometer   | FI21/07010   | 24 <sup>th</sup> , May, 2024  | SGS Fimko Ltd. 0598   | SGS Fimko Ltd. 0598  | 31st, December, 2028                                     | /  |



<sup>3</sup> for devices with AIMDD/MDD certificate(s) the identification should be as in the certificate, and only if the certificate has a generic scope it should be as defined above)